



ELISA Technologies, Inc.

elisa-tek.com



PAGE 1 OF 1 - REPORT 2407189

REPORT DATE: July 31, 2024

REPORT TO: *Adoboloco*

SUBJECT: *Sample for Gluten Analysis*

RECEIVED: *July 26, 2024; USPS PRIORITY MAIL TRK# 9405 5111 0550 1575 4372 26*

ANALYTICAL RESULTS: RESULTS APPLY ONLY TO THE SAMPLES AS RECEIVED.

CONTROL #	SAMPLE DESCRIPTION	ANALYSIS IDENTIFICATION	Gluten
2407189-1	<i>BEST USE BY - 07/19/27 - B01 HBBQ-GUAV</i>	<i>500821 Quantitative Gluten ELISA</i>	Not Detected

NOTE: The ALLER-TEK® Gluten assay is a sandwich ELISA for the quantitative analysis of gluten in raw materials and finished food products. Results are expressed in ppm gluten with a quantifiable range of 2.5-80 ppm; it detects the gliadin and glutenin fractions of gluten. A result of "not detected" indicates that gluten was not detected in the sample (<2.5 ppm). Recovery may be reduced in samples that have been hydrolyzed.

The ALLER-TEK® Gluten ELISA has been validated and certified with license no. 081202 as a Performance Tested Method by the AOAC Research Institute.

Information in italics has been supplied by the customer.

Respectfully Submitted for ELISA Technologies, Inc.

By *Tom d'Angelo*
Analyst

By *JM*
Justin Bickford, Scientific Director

End of Report

THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN ITS ENTIRETY AND WITH THE WRITTEN PERMISSION OF ELISA TECHNOLOGIES, INC.

2501 NW 66th Court, Gainesville, Florida 32653, USA
+1 352 337 3929 info@elisa-tek.com